

Updating the limit lines in your DTX CableAnalyzer

Keeping up to date with the latest draft limit lines is important.

- ISO/IEC 11801 25N1513 Class E_A Permanent Link updated
- ACR-N and PS ACR-N added to TIA Cat 6A Autotest – reports as “i”, information only.
- Bad patch cord warning threshold on 500 MHz Channel test corrected (was too sensitive)

Please click on the link below to download the update file:

<http://www.flukenetworks.com/FNet/en-us/findit?Document=9823410>

With the main unit of the DTX CableAnalyzer connected to your computer and LinkWare 5.0 open:

1. Click on Utilities > DTX CableAnalyzer > Modify DTX Test Limits...
2. Click on To DTX and when prompted to, select the file Limits1.34.limits file and click OK.
3. It will take approx. 20 seconds to do. Click OK when SUCCESS appears. Done.